



# California Medical Device Recall Information



## Recall Name

### Cook Medical Recalls Select Lots of Central Venous Catheters and Pressure Monitoring Sets and Trays Due to a Potential Defect

Recall Date	Product Description	Recalling Firm	Recall Reason
01/06/16	<ul style="list-style-type: none"><li>• Single Lumen Central Venous Catheter Sets and Trays</li><li>• Single Lumen Pressure Monitoring Sets</li><li>• Femoral Artery Pressure Monitoring Catheter Sets and Trays</li><li>• Radial Artery Pressure Monitoring Catheter Sets and Trays</li></ul>	<b>Cook Medical, Inc.</b> Bloomington, IN	<i>Potential for tip fracture or separation which may require medical intervention.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	<a href="#">Affected Product Catalog Numbers and Lot Numbers</a>	<b>CA</b> , nationwide	Distributed between:  April 24, 2015 and October 23, 2015

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm485898.htm>